कार्यालय, नगर पालिक निगम, कोरबा (छ.ग.)

फा.क./स्टोर/2019/1975 कोरबा, दिनांक 29.08.2019

टेंडर नं. 55389
ई–निविदा आमंत्रण सूचना

नगर पालिक निगम कोरबा द्वारा शासकीय अर्द्धशासकीय विभागों में कार्यरत व्यवसाय से संबंधित अनुभवी प्रदायकर्ता/फर्मों/निर्माता कंपनी से (निविदा खुलने के दिनांक तक समस्त संशोधनों के साथ) शेड्युल अनुसार आयटम दरों पर ई–निविदाएं आमंत्रित की जाती है :–

क	कार्य का नाम	प्राक्कलन राशि में	कार्यावधि	अनामत राशि	निविदा खुलने की तिथि
1.	ब्रीथ एनालाईजर (एल्को मीटर) प्रदाय करने बाबत	840000/–	2 माह	8500/–	19.09.2019 से .....

शर्तें :–

1. निविदा दरें आईटम दर पर उल्लेख किया जाए।
2. निविदाकारों द्वारा प्रस्तुत दस्तावेजों का निविदा प्रपत्र में उल्लेखित शर्तों के अनुरूप परीक्षण किया जावेगा। परीक्षण उपरांत पात्र निविदाकारों के ही निविदा को खोला जावेगा।
3. अनुभवी निविदाकार द्वारा नगर पालिक निगम (साडा) के किसी ठेके के कार्य में अनुबंध के अनुरूप कार्य ना किया गया हो अथवा नगर पालिक निगम (साडा) के हित के विरूद्ध कार्य किया गया हो उन्हे निविदा भरने की पात्रता नही होगी।
4. निविदा हेतु इच्छुक निविदाकार को यदि साझेदारी फर्म हो तो उसका प्रमाण पत्र (पार्टनरशीप डीड) की सत्य प्रतिलिपि, पूर्व में इस प्रकार की किए गए कार्यों की प्रमाणित सूची, पिछले वित्तीय वर्ष का आय प्रमाण पत्र/पेन नम्बर पत्र एवं जी.एस.टी. पंजीयन प्रमाण पत्र एवं धरोहर राशि आनलाईन जमा करना होगा एवं जमा की स्केन कॉपी वेबसाईट में अपलोड करना अनिवार्य होगा।
5. सम्पूर्ण किये गये कार्यों के लिये 5 प्रतिशत सुरक्षा राशि देयको के साथ 01 वर्ष के लिये रोकी जावेगी।
6. निविदा के संबंध में किसी प्रकार की विवाद होने की स्थिति में आयुक्त का निर्णय अंतिम एवं सर्वमान्य होगा।
7. सशर्त निविदा स्वीकार नहीं की जावेगीं।
8. बिना कारण बताये निविदा स्वीकृत/अस्वीकृत या पूर्ण रूप से अमान्य करने अथवा स्थगित करने का अधिकार आयुक्त,न.पा.नि.कोरबा को रहेगा।
9. निविदा उन्हीं निविदाकार को दिया जावेगा जिन्हें इस प्रकार के कार्य का अनुभव हो।
10. कार्य हेतु ज्वाइंटवेन्चर मान्य नहीं होगा अर्थात पंजीकृत निविदाकार का अनुभव ही मान्य किया जावेगा।
11. निविदा के संबंध में किसी प्रकार की विवाद होने की स्थिति में आयुक्त का निर्णय अंतिम एवं सर्वमान्य होगा।
12. निविदा स्वीकृत करने अथवा अस्वीकृत करने का अधिकार नगर पालिक निगम के पास सुरक्षित रहेगा।
13. अन्य सूचनाएं नगर पालिक निगम, साकेत भवन रामपुर कोरबा कार्यालय में प्राप्त की जा सकती है।

कार्यपालन अभियन्ता
नगर पालिक निगम
कोरबा (छत्तीसगढ़)

Date-Time Detail(s)

Seq. No.	Nagar Nigam Stage	Contractor Stage	Start		Expiry		Remarks
			Date	Time	Date	Time	
1	Release Tender		29.08.19	10.30	29.08.19	10.30	Release Tender
2		Bid Submission	29.08.19	10.30	18.09.19	17.30	
3	Tender Open		19.09.19	10.30	--	--	

EXECUTIVE ENGINEER
KORBA (CG)
For, Commissioner Korba (C.G.)

Office of The Municipal Corporation, Korba (C.G.)

Supply of Breath Analyser (Alco Meter)

No	Description	Unit	Qty	Rate	Amount
1	Breath Alcohol Analyzer **A8080P** MSIPL Make with Inbuilt Thermal Printer all Accessories with Indian Carry Bag as per attached technical Data in Estimate or as per Engineer -in-charge.	No.	20		

Sub Engineer Mnucipal Corporation Korba (C.G.)	Asstt. Engineer Mnucipal Corporation Korba (C.G.)	Exe. Engineer Mnucipal Corporation Korba (C.G.)

A8080P - The safe Breath Analyzer.

Technical data		
Device:		
Measuring modes:	Active and passive	
Measuring range:	500 mg/dL	
Accuracy:		
Max. Measuring error, in relation to the ethanol standard:		
Range 0 to100 mg/dL	± 5 mg/dL	
Range 1 to 200 mg/dL	± 5 %	of measured value
Range 201upto 500 mg/dL	± 10 %	of measured value
Standard Deviation	0.008‰	
Ambient conditions:		
Temperature range – operation:	-10 °C to +50°C / +23F to +122F	
Temperature range – storage:	-20°C to +60°C / -4F to +140F	
Optimum storage temperature (sensor):	0°C to +25°C / +32F to +77F	
Ambient pressure:	600 to 1400 hPa	
Ambient humidity:	20 to 98 % r.h.	
Minimum expiration volume:	800 mL or depending on requirement	
Batteries:	Inbuilt Li-on Battery.	
Number of measurements per battery pack:	more than 500 tests, depending on the batteries and the ambient conditions	
Device interface:	USB Interface. Data can be transferred to a PC or printer.	
Dimensions (L x W x H):	7inches X 3inches X 1.5 inches.	
Weight:	350 g incl. Batteries	
Miscellaneous:	Ready to measure after switching on:	approx. 5 sec.
	Ready to measure after 0 ‰ measurement:	approx. 2 sec.
	Ready to measure after 1 ‰ measurement:	approx. 5 sec.
Sensor type:	E600 electrochemical measuring cell	
Cross sensitivities:	negligible for other substances, e.g. atcetone, CO etc.	
Measurement units:	all units used world wide configurable	
Operation:	Menu-driven with clear text information	
Input:	QWERTY key pad.	
Power-saving mode:	Device switches off automatically after 2.5 minutes of non-operation, the background lighting switches off after 15 seconds	
Display:	2.4 ich color TFT LCD,fully graphic.	
Data memory:	Stores more than 40,000 measured values with date and time	
Software:	Optional PC software for statistical analysis, data transfer and logging	
Calibration:	Every 6 months	
Mouth Pieces:	With inbuilt Valve to prevent reverse breathing.	
Printer:	Built in thermal Printer as Standard.	
Carry Case:	Local Indian Carry Case	